

Dynamic Navigation in Dental Implant Surgery: A Contemporary Review

Abhishek Gole¹, Shrirang Sangale², Bhushan Bangar³, Susheen Gajare⁴, Shashi Pati⁵, Yogesh Nagargoje⁶, Monika Bageshwar⁷

¹PG Student, ²Lecturer, ³Professor, ⁴Professor & HOD, ⁵Professor, ⁶Lecturer, ⁷PG Student
Dept of Prosthodontic, MIDSAR Dental College, Latur.

Abstract:

Dynamic navigation (DN) for dental implant surgery integrates cone-beam computed tomography (CBCT), optical tracking, and dedicated software to guide implant osteotomy and placement in real time [1]. DN has been shown to reduce angular and linear deviations compared with freehand placement and achieves accuracy similar to static guided surgery, while maintaining greater intra operative flexibility [1, 2, 3]. This review summarizes the principles, workflow, clinical applications, advantages, limitations, and future directions of DN in implant dentistry.

Keywords: dynamic navigation, dental implants, computer-assisted surgery, implant accuracy, CBCT, guided implant surgery

Corresponding Author: Dr. Abhishek Gole, PG Student, Dept of Prosthodontic, MIDSAR Dental College, Latur. Email id.: dranandamb@gmail.com

INTRODUCTION:

Accurate three-dimensional positioning of dental implants is critical for avoiding anatomic complications, optimizing esthetics, and ensuring biomechanically favorable prosthetic rehabilitation [1,4]. Traditional freehand techniques rely on visual and tactile cues and are associated with higher rates of positional error, which may translate into nerve injury, sinus perforation, off-axis loading, and increased prosthetic complexity [1,4].

Common complications related to inaccurate implant positioning include:

- Damage to the inferior alveolar nerve
- Floor of mouth hematoma
- Damage to adjacent roots
- Sinus infections secondary to inadvertent sinus perforations
- Fractured implants due to off-axis loading
- Periimplantitis due to food impaction and off-axis loading

- Poor esthetics secondary to thin buccal, labial bone, and soft tissue
- Interproximal bone loss secondary to placing implants too close to adjacent teeth and implants
- Increased prosthetic complexity and cost

Computer-assisted implant surgery (CAIS) was introduced to address these limitations and now includes static guides and dynamic navigation systems [1,4]. DN provides real-time feedback on drill position, angulation, and depth relative to a virtual treatment plan, and is increasingly adopted in office-based implant practice [1,2].

Classification of Implant Navigation

Implant navigation can be broadly classified into static and dynamic systems [1,4].

Static Navigation (Guided Surgery)

Static navigation, also referred to as computer-guided implant surgery or guided surgery, uses computer-aided design and computer-aided manufacturing (CAD/CAM) surgical templates

based on digital planning of implant position [1]. These templates are fabricated prior to surgery and transfer a preoperative plan to the patient intraoperatively but do not allow real-time modification of implant position or size.

Several types of static surgical guides exist:

- Stone cast-based guides: Basic guides that aid in ensuring appropriate restorable position of the implant but do not take into consideration bone morphology.
- Tooth-supported guides: Fabricated from digital models and seated on remaining dentition, offering good stability and accuracy.
- Tissue-supported guides: Seated on soft tissue or mucosa, more easily displaced than tooth-supported guides.
- Bone-supported guides: Require flap elevation and may offer the highest stability but require additional surgical steps.

Static full-guided, half-guided, and pilot-guided templates demonstrate higher accuracy than non-guided freehand surgery, but their precision depends on CBCT quality, guide fit, sleeve tolerance, and support type [1,4]. DN attempts to retain the accuracy of CAIS while overcoming some of these constraints [1,4].

Dynamic Navigation

DN uses optical tracking and software to display a virtual drill and implant in real time on a monitor, allowing intraoperative adjustments without physical guides. This real-time feedback permits modification of implant position, angulation, length, and diameter during surgery, providing greater flexibility than static systems [1,2].

Principles of Dynamic Navigation Technology Optical Tracking Systems

Most contemporary DN systems used in dentistry employ passive optical tracking [1,2]. An infrared light source illuminates patterned or spherical reflective markers attached to the patient and to the surgical handpiece; a stereo camera system detects these markers and the software reconstructs the spatial position of the drill relative to the patient's CBCT data set in real time [1, 2].




Figure 1: Dynamic navigation system setup showing stereo camera, LED light source, patient tracker, and handpiece tracker positioned for real-time surgical guidance. The overhead camera captures reflective markers to track both patient position and surgical instruments in relation to the planned CBCT coordinate system.

Two types of optical motion tracking systems exist:

1. **Active tracking systems:** Arrays emit infrared light that is tracked by stereo cameras.
2. **Passive tracking systems:** Arrays use reflective spheres or patterned markers to reflect infrared light emitted from a light source back to a camera.

The patient and drill must remain within the line of sight of the tracking camera. Light is projected from a light-emitting diode (LED) light source above the patient, reflects off tracking arrays attached to the patient and the surgical instrument, and is captured by stereo cameras. The DN system then calculates the position of the patient and instruments relative to the presurgical plan in real time [1].

Workflow Overview

The core steps of DN are:

1. **Image Acquisition:** Obtaining a 3D CBCT dataset including the surgical field and fiducial markers.
2. **Virtual Planning:** Planning of implant position, angulation, and depth within dedicated navigation software.
3. **Calibration:** Calibration of the handpiece, drills, and probe to the tracking system to determine the geometric relationship between tracking markers and the instrument tip.

4. Registration: Registration of the patient tracker to the fiducials, linking the physical patient to the virtual CBCT coordinate system.
5. Guided Surgery: Performing osteotomy and implant placement while monitoring real-time position feedback on the navigation screen.

The system displays a virtual drill superimposed on multiplanar views and 3D reconstructions, with numerical readouts of angular deviation and depth relative to the planned implant [1,2].

Clinical Workflow and Technique

Fiducial Strategies

Fiducial markers are essential for linking the patient's physical anatomy to the virtual CBCT coordinate system. Strategies differ for dentate and edentulous patients.

4.1.1 Dentate Patient Fiducial

In dentate patients, a thermoformed or thermoplastic fiducial clip is adapted over several teeth to create a rigid, reproducible support for the fiducial markers and patient tracker[1].



Figure 2: Figure 2: Fiducial clip positioned on maxillary anterior teeth demonstrating proper thermoplastic adaptation with buccal tracker arm. The clip must be firmly seated without rocking and positioned to minimize optical interference from surgeon's hands and instruments during the procedure.

Preparation of the fiducial clip:

The fiducial clip is placed in a hot water bath at a temperature of 60°C-71°C (140°F-160°F) for approximately 3-5 minutes until the thermoplastic becomes clear[1]. The clip is then cooled for approximately 1 minute to reach a surface temperature below 40°C (104°F) before insertion in the mouth[1].

Seating technique:

The fiducial clip should be seated on three teeth, ensuring equal distance on the buccal and lingual sides, with the tracker arm positioned on the buccal side[1]. Vertical pressure is applied until the plastic surface cannot advance further. An adequate impression is confirmed, and the clip is removed without any rocking motion and placed in a cold water bath[1].

Clinical considerations:

The fiducial clip must be:

- Firmly supported by teeth without any mobility when seated
- Placed on teeth that are not mobile, do not serve as pontics on a bridge, and are free of orthodontic wires
- Positioned to minimize optical interference by the surgeon's or assistant's hands and instruments
- Tried again in the mouth to confirm accuracy and ensure there is no impingement of soft tissue

If clinical crowns are short or teeth lack undercuts, composite resin may be added to the buccal and occlusal surfaces of associated teeth to help create immobile fiducial clip insertion[1].

Edentulous Patient Fiducial

An edentulous patient case requires edentulous fiducials (small self-tapping screws) to be placed in the patient's bone to facilitate registration in the CT scan[1].



Figure 3: Figure 3: Distribution of edentulous fiducial screws in a partially edentulous arch. Minimum of 4 screws (1.5 mm diameter, 4-5 mm length) positioned throughout the arch, avoiding planned implant sites, inferior alveolar nerve canal, and tracker plate location. This configuration ensures accurate registration of patient anatomy to CBCT coordinate system.

Fiducial screw specifications:

Edentulous fiducial screws should be:

- Diameter: 1.5 mm
- Length: 4-5 mm (4 mm recommended in posterior mandible or areas of dense cortical bone; 5 mm or greater in maxilla or regions of immature, soft, or grafted bone)
- Self-drilling and self-tapping
- Low profile and stable

When placing fiducials in the mandible, short 4-mm screws should be used to avoid damage to the inferior alveolar nerve[1]. The inferior alveolar nerve and infraorbital nerve must be considered and avoided when placing fiducials[1].

Placement strategy:

A minimum of 4 fiducials should be placed and spread throughout the arch, leaving room for an edentulous fiducial plate to be inserted at the time of surgery[1]. Edentulous fiducials must be placed in

the arch where implants will be placed. If implants are to be placed in both the maxilla and mandible, fiducials must be placed in both arches. If vertical bone reduction is anticipated, fiducials must be placed apical to the area of proposed bone reduction [1].

Image Acquisition and Software Planning CBCT Acquisition

Image acquisition includes obtaining 3D files, usually a CBCT in Digital Imaging and Communications in Medicine (DICOM) format (.dicom)[1]. The field of view of the CBCT should include the surgical site and all fiducials[1]. The scan is obtained with the plane of occlusion of the implant site parallel to the floor [1].

Soft tissue visualization:

An important point related to CBCT acquisition that is often overlooked is the separation of soft tissues while taking the image[1]. For dental implant planning purposes, a cotton roll or radiolucent material placed between the dentition and the buccal/labial mucosa creates an air contrast zone, allowing visualization of the soft tissue in the region of the free gingival margin on the CBCT[1].

Dual Scan Technique

Dual scan is the term used when a dental appliance, such as a set of dentures, is superimposed over a patient's CT scan[1]. If a dual scan technique is utilized, at least five 2-mm fiducials should be applied to the denture. A high-resolution CT scan is obtained of the denture on its own and then a separate CT scan is obtained with the denture in the patient's mouth, ensuring not to disturb the fiducials on the patient and denture[1].

Intraoral Scanner Integration

An alternative to dual scan is the use of an intraoral scanner (IOS)[1]. An IOS provides a 3D surface image of the patient's dentition and occlusion. These are not volumetric images; IOS images are a surface. IOS images have a high degree of accuracy for single and quadrant impressions; when full arches are scanned, the accuracy decreases[1].

The implant team may wish to obtain IOS of the patient before teeth are extracted. If the occlusion is not going to be changed, these images can be saved for later use for planning of ideal implant position and provisional fabrication [1].

Virtual Planning

Once images are acquired and stored, they are loaded into treatment planning software. Several software packages are available, but key features should include the ability to:




Figure 4: Figure 4: CBCT navigation planning interface showing multiplanar views (axial, sagittal, coronal) with superimposed 3D surface scan of teeth. Virtual implant crowns are positioned in proper occlusal relationship; with underlying virtual implant axes demonstrating prosthetically-driven planning that ensure ideal emergence profiles and biomechanical positioning.

- Import and export generic file formats (.dicom and .stl)
- Superimpose 3D files
- Perform dual scan .dicom superimposition
- Export images in a common coordinate system as individual or merged items

When clean .stl images are superimposed on CBCT data, the combined images allow the implant team to

plan with osseous, dental, and soft tissue structures clearly visible along with the patient's occlusion[1].

Restoratively-Driven Planning

When starting to plan on the DN software, a panoramic curve for the arch requiring implants is developed on the axial plane of the patient's scan[1]. On the mandible, the inferior alveolar nerve also can be identified and marked[1].

Merger of the patient's scan and the IOS image or denture scan is performed, ensuring multiple areas of coordination between the images for accuracy of the merger[1].

Critical principle: The planning of implants should be restoratively driven. This starts with evaluating the occlusion and placing the restorative envelope of the virtual teeth in the proper occlusal position using virtual implant crowns available in the DN software[1]. Alternatively, a separate prosthetic software can be used to plan the restorations, which are then exported from the prosthetic software and imported as a .stl file into the DN software[1].

Once the implant crown is finalized, the virtual implants should be properly aligned below the virtual crowns for ideal emergence into the prosthetic space[1]. The DN software allows design of a generic implant or previously specified implant, with customization of implant platform diameter, apex diameter, length, and abutment height and angle[1]. Additional tools in the DN software allow mirroring to align implants across an arch and paralleling of adjacent implants[1].

Calibration

The instruments to be tracked by the system during surgery must be calibrated[1]. The geometry of the tracking arrays relative to the instrument being used must be determined by the tracking system. The assembled parts must be placed in front of the stereo cameras so the software can "learn" their geometry[1].

The instruments to be calibrated include:

1. Contra-angle handpiece
2. Straight handpiece
3. Probe tool

The calibration of instrumentation occurs approximately 60–80 cm from the camera[1]. The

contra-angle handpiece along with the handpiece tracker is assembled and calibrated by rotating the handpiece such that the camera can locate and identify the patterns on the handpiece tracker[1].

After calibration of the handpiece, a contra-angle handpiece chuck calibration is performed[1]. The handpiece is attached to the chuck and the drill motor is run at 10–20 revolutions per minute over the camera to calibrate the chuck plate to the handpiece[1]. A Go Plate (X-Nav Technologies, LLC, Lansdale, Pennsylvania) and probe are calibrated by placing the probe in the pivot hole of the Go Plate[1]. An implant drill bit is placed on the handpiece and the drill bit is placed on the Go Plate perpendicular to the center target[1]. The drill length is then verified by the DN system[1].

Registration:

The DN system must also be "taught" the geometry of the patient tracking array relative to the fiducials and thus the planned implants. This process is called registration[1].

Dentate Patient Registration:

For the dentate patient, the fiducial clip attached to a patient tracker arm and patient tracker is registered automatically by the system at the time of calibration[1].

Edentulous Patient Registration:

In the edentulous patient, an edentulous patient calibration probe is calibrated[1]. Then, the edentulous tracker plate is placed on the bone of the patient underneath a subperiosteal flap in an area of bone where there are no edentulous fiducial screws[1]. The tracker plate is attached to a patient tracker arm and patient tracker[1].

The patient tracker and the edentulous fiducial screws are then registered to the DN system by touching the screws (fiducials) with the probe as the system tracks them[1].

Verification of Calibration Accuracy:

The calibration accuracy is verified between the fiducials and the drill[1]. The drill bit is placed on three fiducial spheres on the fiducial clip for the dentate patient or the edentulous fiducial screws[1].

The doctor looks at the two-dimensional (2D) views for accuracy data in green colors[1]. If all three fiducials have green indicators, the system calibration is within 200 micrometers[1]. This step is not performed with edentulous patients[1].

Preoperative System Check:

Prior to the start of surgery and after every drill bit is changed, a "system check" is performed by the doctor[1]. This step ensures the instruments are calibrated and the system is properly registered to the patient[1].

Intraoperative Navigation Surgery:

It is important to always confirm the accuracy of the tracking system by performing frequent system checks[1]. Anatomical landmarks on the patient are touched with the instruments. The doctor then visually confirms that the radiographic landmarks on the screen are exactly correlating[1].

Optimal landmarks are adjacent teeth or bony landmarks close to the planned implant site or fiducial markers on edentulous patients[1]. The operator looks at the screen as the drill is positioned over the surgical site[1]. The navigation system screen allows viewing of a virtual drill with demonstration of:

- Depth in tenths of a millimeter
- Angular deviation of the drill bit axis from the planned implant axis to the tenths of a degree
- The implant timing

The tip of the drill, indicated by a blue dot, is positioned over the target to indicate ideal planned platform position[1]. The top of the drill, a small circle, is then centered over the blue dot to indicate ideal planned angle[1]. Depth is indicated by color: yellow, green, and red[1]. The planned depth is always at the 45 degree position on the target[1].

The surgical assistant is in charge of suctioning and looking into the surgical field to notify the surgeon of any irregularities such as lack of irrigation or grossly off-positioned drill placement[1].

As implant drilling occurs, the depth indicator changes in color from green to yellow when the drill is 0.5 mm from the targeted depth[1]. The yellow turns to red, indicating when to stop the depth of the osteotomy[1].




Figure 5: Figure 6: Intraoperative navigation monitor display showing real-time surgical guidance during implant osteotomy. The screen displays multiplanar CBCT views with virtual drill overlay, 3D reconstruction of planned implant position, angular deviation readout (in tenths of degrees), and color-coded depth indicator (green-yellow-red) showing proximity to target depth. This real-time feedback enables precise control and continuous accuracy verification throughout the procedure.

Intraoperative flexibility: During the implant surgery, the implant size, width, type, and location can be adjusted based on intraoperative factors deemed necessary for a stable and appropriately restorable implant[1]. This flexibility represents a significant advantage over static guided surgery[1].

Accuracy and Clinical Outcomes:

Multiple in vitro and clinical studies have evaluated the accuracy of DN compared with static guides and freehand placement[1,2,3].

Mean Deviation Measurements

A systematic review and meta-analysis of dynamic systems reported:

- Mean global platform deviation: approximately 1.0 mm
- Mean apical deviation: approximately 1.3 mm
- Mean angular deviation: 3.6–4.1 degrees
- No significant difference between model and clinical studies or between jaws

Prospective clinical data indicate that both DN and static guides achieve significantly lower platform, apical, and angular deviations than freehand placement[1,2,3]. Freehand placement shows a mean angular deviation of 6.50–9.92 degrees, whereas DN-placed implants demonstrate a mean angular deviation of 2.97–3.6 degrees[1,2].

Comparative Accuracy

Comparison of different implant placement methods:

Placement Method	Mean Angular Deviation (degrees)	Accuracy
Freehand (mucosal-borne guide)	9.92 ± 6.01	Lowest
Freehand (general)	6.50 ± 4.21	Low
Dynamic Navigation	2.97 ± 2.08	High
Static Guided Surgery	2.71 ± 1.36	High

Table 1: Comparison of mean angular deviations for different implant placement methods




Figure 6: Figure 5: Bar chart comparing mean angular deviations (in degrees) across different implant placement methods. Freehand techniques demonstrate significantly higher deviations, with mucosal-borne guides showing 9.92±6.01° and

general freehand $6.50 \pm 4.21^\circ$, whereas both dynamic navigation ($2.97 \pm 2.08^\circ$) and static guided surgery ($2.71 \pm 1.36^\circ$) achieve substantially improved accuracy. Error bars represent standard deviation around each mean value.

Similar linear accuracy is achieved between DN and static guides, with slightly higher angular accuracy in some static systems[1,2]. Nevertheless, deviations obtained with DN remain well within clinically acceptable ranges and comparable to high-quality static guided surgery[1,2].

Clinical Applications

Implant survival rates with DN appear similar to those of conventional and static guided techniques[1,2,3]. DN has been successfully applied to complex situations including:

- Atrophic jaws with limited bone volume
- Immediate implant placement
- Pterygoid implants for extended cantilevers
- Full-arch rehabilitations and multiple implant cases
- Narrow interdental spaces with limited guide placement
- Patients with limited mouth opening

Advantages of Dynamic Navigation

Key advantages of DN over freehand and static guided surgery include:

Improved Accuracy and Safety

DN reduces angular and linear deviations relative to freehand placement, reducing the risk of:

- Nerve injury (inferior alveolar nerve, infraorbital nerve)
- Sinus perforation
- Cortical plate perforation
- Adjacent root damage
- Prosthetic misalignment and esthetic complications

Any form of computer-assisted surgery (CAS) is statistically more accurate and precise than freehand placement because it overcomes the inherent inaccuracy of human vision [1,2].

Real-Time Feedback and Verification

Real-time verification of drill trajectory and depth during every step of osteotomy preparation and implant insertion allows the surgeon to monitor position accuracy continuously[1,2]. Unlike static guides, where a mis-seated splint may result in significant gross error of the entire implant surgery, DN allows continuous verification of accuracy throughout the procedure[1].

Intraoperative Flexibility:

The ability to modify implant position, length, diameter, and even system during surgery without fabricating a new guide is a critical advantage over static templates[1]. This flexibility addresses the limitations of static guides, where changing implant position requires fabrication of a new guide or abandonment of the guided approach[1].

Minimally Invasive Approach:

DN facilitates flapless or minimally invasive approaches, potentially reducing:

- Intraoperative trauma
- Postoperative pain
- Healing time
- Tissue morbidity

DN allows surgeons the confidence to know implant placement is appropriately in bone without having to open a flap, thus minimizing trauma to the patient[1].

Ergonomic Benefits:

Ergonomically, DN allows the surgeon to look at the screen more so than inside the mouth, decreasing the need to bend the back or neck for a prolonged period[1]. This improved posture reduces surgeon fatigue and strain[1].

DN also allows the surgeon to perform the osteotomy and place the implant with limited direct visualization in the mouth in patients with:

- Limited mouth opening
- Difficult visualization in posterior regions
- Narrow interdental spaces that prohibit appropriate guidance tubes with static guides

Same-Day Treatment Protocol:

Implant surgeons are able to evaluate a patient, scan the patient, plan the implant position, and perform the implant surgery in the same day without the delay or cost of fabrication of a static surgical guide

stent[1]. This streamlined workflow improves efficiency and patient satisfaction[1].

Patient-Reported Outcomes:

Patient-reported outcomes suggest that DN-assisted surgery may be perceived as more comfortable and less traumatic. Patients often appreciate being able to visualize the procedure on the screen and feel reassured by real-time monitoring[1].

Limitations and Disadvantages:

Despite its benefits, DN presents several challenges and limitations that clinicians should consider before implementation.

8.1 Financial Investment

The implementation of DN requires significant investment for the dental implant surgeon:

1. Capital costs: DN system hardware and software licenses
2. Imaging infrastructure: CBCT and intraoral scanning equipment
3. Per-case costs: Fiducial clips, markers, plates, and tracking components
4. Ongoing expenses: Software maintenance and system support

These costs may limit accessibility for smaller practices or those with limited capital budgets[1,2].

Learning Curve:

Those surgeons with limited experience with technology and virtual image processing may find it difficult to transition to a different modality of practice[1]. There is a learning curve with the application of a new technology for all levels of technological comfort[1].

The learning curve for DN systems was evaluated, showing that surgeons become statistically equivalent and proficient after 10–20 implants placed with the system[1]. During the learning phase, efficiency may be reduced and case selection limited[1].

In addition, restorative dentists will require training to be comfortable with the workflow implemented by the implant surgeon[1]. Interdisciplinary communication and coordination are essential for optimal outcomes[1].

Additional Surgical Steps for Edentulous Patients:

The current FDA-approved systems for edentulous patients require the additional surgery of the

placement of fiducial screws and tracking plates[1]. This adds:

- Increased surgical time and cost
- Increased invasiveness for edentulous patients
- Additional patient morbidity from screw placement
- Need for coordinated scheduling of fiducial placement and implant placement

However, this obstacle will soon be replaced with a fiducial-free method, where the patient's anatomy will take the place of the screws[1].

Line-of-Sight Requirement:

Dependence on line-of-sight between the camera and tracking arrays represents a technical limitation[1,2]. Obstruction by hands, instruments, or soft tissues can temporarily interrupt tracking and require repositioning or re-registration[1]. In complex surgical cases or those with limited mouth opening, maintaining clear line-of-sight can be challenging[1].

Calibration and Registration Complexity:

Requirement for meticulous calibration and frequent system checks is essential; errors in these stages can compromise accuracy as much as mis-seated static guides[1,2]. The multi-step calibration and registration process requires:

- Careful attention to detail
- Understanding of system mechanics
- Systematic verification steps
- Knowledge of troubleshooting and error correction

Any deviation in calibration accuracy will propagate through the surgical procedure and compromise implant positioning[1].

Technological Dependence:

DN systems are complex technological devices that may experience hardware or software malfunctions[1,2]. Dependence on technology introduces the potential for system failures that may not be immediately apparent and could compromise accuracy[1]. Surgeons must have contingency plans and be prepared to revert to freehand or static guided techniques if system failure occurs[1].

Future Directions:

Integration of DN with emerging technologies represents an active and promising area of research and development[1,2,3].

Fiducial-Free Registration

Future developments include fiducial-free registration using anatomical landmarks [1]. Rather than requiring placement of fiducial screws in edentulous patients or thermoformed clips in dentate patients, the surgeon will select specific anatomical points on the CBCT during planning. After the patient tracker is placed, the patient will be registered by touching these points with the calibrated probe[1]. This approach will reduce invasiveness, operative time, and cost[1].

Robotic-Assisted Dynamic Navigation:

Robotic CAIS systems offer the potential for even higher positional stability and accuracy, particularly for long-span and extra-maxillary implants[1,3]. Integration of DN with robotic technology may allow autonomous or semi-autonomous execution of the surgical plan while maintaining real-time monitoring and intraoperative flexibility[1].

Artificial Intelligence and Machine Learning:

Integration of DN with artificial intelligence (AI) and machine learning offers several potential applications:

- Automated implant position optimization based on bone quality, anatomy, and occlusal considerations
- Predictive modeling of long-term outcomes based on implant position and angulation
- Real-time image guidance and augmented reality visualization
- Automated detection and correction of navigation system drift
- Personalized surgical planning based on individual anatomic variation

Augmented Reality Visualization

Augmented reality (AR) overlays of the virtual implant and surgical plan directly into the surgeon's field of view (using AR glasses or headsets) may eliminate the need to focus on a separate monitor and further improve ergonomics and surgical efficiency[3].

Cost Reduction and Technology Maturation:

As DN systems hardware and software mature, several disadvantages will diminish:

- Reduction in equipment costs through competition and economies of scale
- Simplified calibration and registration workflows
- Improved reliability and reduced downtime
- Wider accessibility across practice settings
- Standardization of techniques and better training resources

Clinical Research Needs

Further clinical research is needed to:

- Evaluate long-term survival rates, marginal bone behavior, and biological complications specific to DN workflows
- Assess patient-centered outcomes such as pain, satisfaction, quality of life, and return to function
- Perform cost-effectiveness analyses comparing DN with static guides and freehand protocols across different practice settings
- Develop evidence-based guidelines for case selection and patient counseling
- Establish standardized accuracy metrics and quality assurance protocols
- Compare DN with emerging robotic and AI-assisted systems

CONCLUSION:

The natural progression from analog 2D imaging and diagnostics to digital 3D imaging and diagnostics has led to increased understanding of the complex nature of implant surgery and prosthodontics[1]. The increased utilization of these digital 3D diagnostic and therapeutic modalities allows the surgical team to see the limitations of freehand surgery[1].

Computer-assisted surgery allows the implant team to overcome the limitations of human stereo vision and increase the accuracy and precision of implant placement [1]. DN allows the surgeon to implement digital implant treatment plans in an efficient fashion[1].

This efficiency and flexibility allow the team to utilize CAIS on every implant in every patient[1]. High-level statistical evidence clearly illustrates the improved accuracy and precision of computer-assisted surgery over freehand surgery[1,2,3].

Dynamic navigation represents a significant evolution in implant surgical technology, offering clinicians a powerful tool to enhance accuracy, minimize complications, and optimize patient outcomes. As technology matures, costs decrease, and evidence accumulates, DN is likely to become increasingly integrated into mainstream implant practice. Continued innovation and research will further refine these systems and expand their clinical applications.

REFERENCE

1. Panchal, N., Mahmood, L., Retana, A., & Emery III, R. (2019). Dynamic Navigation for Dental Implant Surgery. *Oral and Maxillofacial Surgery Clinics of North America*, 31(4), 539–547. <https://doi.org/10.1016/j.coms.2019.08.001>
2. Block, M. S., & Emery, R. W. (2016). Static or dynamic navigation for implant placement—choosing the method of guidance. *Journal of Oral and Maxillofacial Surgery*, 74(2), 269–277. <https://doi.org/10.1016/j.joms.2015.08.023>
3. Emery, R. W., Merritt, S. A., Lank, K., & Block, M. S. (2016). Accuracy of dynamic navigation for dental implant placement—model-based evaluation. *Journal of Oral Implantology*, 42(5), 399–405. <https://doi.org/10.1563/aaaid-joi-d-15-00162>
4. Rungcharassaeng, K., Caruso, J. M., Kan, J. Y., Zimmerman, G., & Kim, J. S. (2015). Accuracy of computer-guided surgery: A comparison of operator experience. *Journal of Prosthetic Dentistry*, 114(3), 407–413. <https://doi.org/10.1016/j.prosdent.2015.04.016>